

LAAnatomically Realistic Simulations of Liver Ablation by Irreversible Electroporation: Impact of Blood Vessels on Ablation Volumes and Undertreatment

Radwan Qasrawi^{1*}, Ph.D. Candidate¹, Louis Silve², Fernando Burdío, PhD³, Ziad Abdeen, PhD⁴, and Antoni Ivorra, PhD^{1,5}

¹ Department of Information and Communication Technologies, Universitat Pompeu Fabra, Barcelona, Spain. ² Ecole Centrale de Lyon, France. ³ Department of Surgery, Hospital del Mar, Barcelona, Spain. Hospital del Mar Medical Research Institute (IMIM), Barcelona, Spain. ⁴ Faculty of Medicine, Al-Quds University, Jerusalem, Palestine.

⁵ Serra Húnter Fellow, Universitat Pompeu Fabra, Barcelona, Spain.

Abstract

Irreversible electroporation is a novel tissue ablation technique which entails delivering intense electrical pulses to target tissue, hence producing fatal defects in the cell membrane. The present study numerically analyzes the potential impact of liver blood vessels on ablation by irreversible electroporation because of their influence on the electric field distribution. An anatomically realistic computer model of the liver and its vasculature within an abdominal section was employed, and blood vessels down to 0.4 mm in diameter were considered. In this model, the electric field distribution was simulated in a large series of scenarios ($n = 576$) corresponding to plausible percutaneous irreversible electroporation treatments by needle electrode pairs. These modeled treatments were relatively superficial (maximum penetration depth of the electrode within the liver = 26 mm) and it was ensured that the electrodes did not penetrate the vessels nor were in contact with them. In terms of total ablation volume, the maximum deviation caused by the presence of the vessels was 6%, which could be considered negligible compared to the impact by other sources of uncertainty. Sublethal field magnitudes were noticed around vessels covering volumes of up to 228 mm³. If in the model the blood was substituted by a liquid with a low electrical conductivity (0.1 S/m) the maximum volume covered by sublethal field magnitudes was 3.7 mm³ and almost no sublethal regions were observable. We conclude that undertreatment around blood vessels may occur in current liver ablation procedures by irreversible electroporation. Infusion of isotonic low conductivity liquids into the liver vasculature could prevent this risk.

Keywords

Irreversible electroporation, liver, conductivity, numerical modeling, blood vessels.

Abbreviations CT, computer tomography; IRE, irreversible electroporation; RF, radiofrequency.

Introduction

Irreversible electroporation (IRE) is a relatively novel non-thermal ablation modality (1). IRE is typically performed by inserting thin electrodes into the target tissue and delivering a number (8 to 100) of short (10 μ s to 100 μ s) high voltage pulses across the electrodes thereby producing field magnitudes in the tissue in an order of 1000 V/cm. These pulsed fields cause a large increase in cell membrane permeability to ions and macromolecules which fatally disrupts cell homeostasis (2).

Ablation of soft tumors by IRE has been demonstrated in a number of clinical studies (3–9). IRE ablation is not based on heating, which gives it two major advantages over other ablation modalities: 1) IRE does not destroy the extracellular matrix, and 2) IRE is insensitive to thermal sinks. These advantages imply that it is safe and physically possible to use IRE to treat tissues which are in close proximity to vital vessels or are in contact with them. The first advantage implies that it is safe to perform IRE ablation in the vicinity of vital vessels because the structural properties of those vessels are preserved (8,9). It is even possible to directly treat arteries without compromising their fluidic function (10). The second advantage implies that it is physically possible to perform IRE ablation in the vicinity of large blood vessels (11,12). In contrast to other ablation modalities based on heating or freezing, the thermal sink effect produced by large blood vessels does not hinder IRE ablation. From this last feature, it may be mistakenly concluded that the presence of blood vessels is irrelevant for IRE ablation. A subjacent objective of the present work was to illustrate that this is not the case.

IRE's effectiveness depends on the temporal features of the sequence of pulses, but for a given set of temporal features (number of pulses, pulse duration, and repetition frequency), it is accepted that local IRE effectiveness depends mainly on the local field magnitude; an electric field threshold for IRE is typically defined for each tissue type and pulse protocol (13). Therefore, any factor that may have an impact on the electric field distribution may have an impact on the ablation. In the case of the liver – which tends to be considered as homogenous in electrical terms – one of those factors may be the contrast in electrical conductivity between the liver parenchyma and the interior of the blood vessels. Indeed, recent numerical studies (14-16) and *in vivo* studies (14,17) have revealed that this sort heterogeneity may have a significant impact on IRE ablation. However, to the best of our knowledge, its impact on liver ablation has been neither observed clinically nor assessed for clinical relevance through numerical studies.

The purpose of the study reported here is to numerically analyze and quantify the potential impact of liver blood vessels on ablation by irreversible electroporation (IRE) due to their influence on the electric field distribution.

Materials and Methods

An anatomically realistic computer model of a male human liver, its blood vessels (down to 0.4 mm in diameter), bones and surrounding tissues was employed in this study. The 3D anatomical geometry was developed by the Korea Institute of Science and Technology Information (18). The geometry was decimated and prepared for simulation using the Graphical Interface for Medical Image Analysis and Simulation (GIMIAS) software platform developed by the Center for Computational Imaging and Simulation

Technology in Biomedicine (CISTIB) at the Universitat Pompeu Fabra (19). The main geometrical features of the modeled liver and its vasculature are indicated in Table 1.

Table 1. Geometrical measurements of the liver and the vessels used in this study				
Blood Vessels	Diameter		Length	Volume
	Min(mm)	Max (mm)	(X,Y,Z)(mm)	(ml)
Left vein	1.4	10.9	95.0, 79.6, 60.6	5.3
Left duct	0.4	4.4	122.3, 46.7, 39.6	0.6
Left artery	0.5	2.9	6.3, 30.8, 19.8	0.6
Middle vein	1.2	13.2	63.0, 80.7, 74.4	5.3
Right vein	0.8	14.2	93.3, 62.5, 121.2	7.3
Right duct	0.5	5.1	33.0, 69.3, 60.4	0.3
Right artery	0.4	3.0	50.3, 22.9, 36.7	0.3
Portal vein	1.2	18.3	130.4, 114.4, 88	9.6
Liver	-	-	230.9, 170.5, 161.8	1514.3

Electroporation across pairs of needle probes was simulated. The two probes of the pair were parallel and aligned so that they would form the opposite sides of an imaginary rectangle. Six arbitrarily chosen locations and four separations (10, 15, 20 and 25 mm) per location were assayed. These locations correspond to probe insertions through the anterior part of the abdomen. Figure 1 shows the anatomic model together with the assayed locations (separation = 25 mm). Each needle probe, with a diameter of 1 mm, consisted of a passive shaft (length = 40 mm) and an active part (i.e. the electrode) with a length of 10, 15, 20, or 25 mm. The penetration depth of the pair was adjusted to ensure a complete and superficial insertion of the active part into the liver with a maximum proximal distance of 1 mm from the surface of the liver. This resulted in a maximum penetration depth of 26 mm within the liver.




Figure 1. 3D geometrical representation of the 6 insertion locations assayed for the electrode pairs (in this case the interelectrode spacing is 25 mm and the electrode length is 25mm).

Electrode locations were carefully checked for ensuring that the electrodes neither penetrated the vessels nor were in contact with them. It was ensured that the minimum distance between the electrodes and any vessel was 0.2 mm. It was also ensured that the probes did not go through bone tissue. If these conditions were not met, the location of the probe pair was randomly modified a few millimeters (<5 mm) maintaining the above constraints regarding parallelism, separation and penetration depth.

For each model (homogeneous and non-homogeneous) three voltages across the two electrodes were assayed: 2000, 2500 and 3000 V. The total number of simulated cases for each model was 288 (6 locations \times 4 separations \times 4 electrode lengths \times 3 voltage magnitudes). The electric field threshold for effective irreversible electroporation was considered to be 700 V/cm (2,15, 20).

The 3D geometry was imported into COMSOL Multiphysics 4.3a (COMSOL AB, Stockholm, Sweden). The entire geometry was meshed using the COMSOL Multiphysics regular refinement method user controlled tetrahedral mesh with a minimum element size of 0.1 mm. The total number of elements was above 2,200,000.

The electric field distribution was computed using the ‘Electric Currents (ec)’ application mode of the AC/DC module of COMSOL. As in previous studies (15,21-24), non-linear behavior was assumed for the hepatic tissue in order to model the increase in tissue electrical conductivity during pulse delivery due to electroporation. This behavior was modeled with the following sigmoid equation describing the dependency of tissue conductivity (σ) of electric field magnitude ($|\vec{E}|$):

$$\sigma = \sigma_0 + [(\sigma_f - \sigma_0) \frac{1}{1 + e^{-\frac{|E| - T}{W}}}] \quad (1)$$

Where σ_0 is the conductivity of the tissue when no field is applied (i.e. initial conductivity), σ_f is the maximum conductivity the tissue can reach due to electroporation (i.e. final conductivity), and T and W are two parameters that describe the morphology of the sigmoid (coarsely, inflection magnitude and slope respectively). The set of values used for the hepatic tissue model was: $\sigma_0 = 0.2$ S/m, $\sigma_f = 0.5$ S/m, $T = 950$ V/cm and $W = 200$ V/cm. These values would correspond to hepatic tumors and have been inferred from experimental data in Prakash et al. (25). The values for healthy (normal) liver tissue would be $\sigma_0 = 0.05$ S/m, $\sigma_f = 0.3$ S/m, $T = 950$ V/cm and $W = 200$ V/cm; also inferred from data in Prakash et al. (25). This set of values was also assayed in this study and the results obtained are reported as supplementary material. We deemed it more realistic to model tumor properties rather than healthy liver properties because IRE treatments mostly occur on tumor tissues (i.e. the electrodes are inserted into the tumors).

The electrical conductivity of the other tissues was modeled as field independent. The conductivities of the vessels (modeled as blood containers), bones and soft tissues filling the anatomy (modeled as a combination of muscle, fat and body fluids) were 0.7 S/m, 0.02 S/m and 0.2 S/m respectively; these values were obtained from experimental data found in Gabriel et al. (26). Conductivity values of the electrodes and the passive shafts were set to 1×10^5 S/m and 1×10^{-5} S/m respectively.

For each case, a simulation was performed assuming the presence of the vasculature ($\sigma_{\text{vessels}} = 0.7$ S/cm) and another one was performed under the assumption that the liver was homogenous ($\sigma_{\text{vessels}} = \sigma_{\text{liver}}$). Both solutions were then compared. First, the

size of the ablation volumes were compared (regions in which $|\vec{E}| > 700$ V/cm). It was then checked whether there were any regions in the ablation volume predicted with the homogeneous assumption ($\sigma_{\text{vessels}} = \sigma_{\text{liver}}$) that were not included in the ablation volume predicted with the heterogeneous assumption ($\sigma_{\text{vessels}} = 0.7$ S/cm). That is, the solution of the homogeneous assumption was considered to be the target volume and it was checked whether the blood vessels caused undertreatment within that volume.

In addition to the above referred systematic study with 288 cases, for graphical illustration, two hypothetical treatment cases were simulated to emphasize the impact of undertreatment around vessels. First, case H1 was simulated in which the presence of a tumor ($20.7 \times 19 \times 10$ mm³) was modeled at a distance of 0.3 mm away from the right hepatic vein (Figure 2a). In this case, the distinctive conductive properties of the liver and of the tumor were modeled according to the tumor conductivity models and parameters described above. The second case (case H2), the impact of blood vessels on the IRE treatment was assessed by inserting the pair electrodes (active part =25mm and passive shaft =50 mm) deep into the liver with a maximum penetration depth of 65mm. The electrodes were located 1 mm from the right hepatic vein and the portal vein which are 10.2 mm and 14.6 mm in diameter respectively, and 3 mm from the middle hepatic vein which is 8.1 mm in diameter.

Furthermore, as it is common to employ more than two electrodes in IRE treatments, additional simulations were performed using three-electrode and four-electrode arrays. For each array model (homogeneous and non-homogeneous) two active length used (20 and 25mm), two interelectrode spacing (20 and 25mm) and three voltages

across the electrodes pairs (2000, 2500 and 3000V). The total number of simulated cases for each array model was 12 (2 separations \times 2 electrode lengths \times 3 voltage magnitudes).

Results

Simulations indicate that undertreatment may appear around vessels

Figure 2a shows the geometry of the modeled hypothetical treatment of a tumor (case H1) and Figures 2b and 2c show a cross section of the corresponding simulated electric field magnitude assuming that the liver is homogeneous ($\sigma_{\text{vessels}} = \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) and that it is heterogeneous ($\sigma_{\text{vessels}} = 0.7 \text{ S/cm} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) respectively. It can be observed that for the homogeneous case (Figure 2b) the treatment volume (regions in which $|\vec{E}| > 700 \text{ V/cm}$) well covers the tumor (voltage across electrodes = 2000 V). However, once the blood vessels are modeled (Figure 2c), the electric field distribution is distorted around them. In particular, it could be noticed that there is a tumor region close to the right hepatic vein (lower vessel) in which the electric field is significantly below 700 V/cm.




Figure 2. X-Y plane plot of electric field magnitude during hypothetical treatment of a tumor. (a) The 3D geometry model. (b) The homogeneous model ($\sigma_{\text{vessels}} = \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) result. (c) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) result. (d) Result when the vessel is filled with a low conductivity fluid ($\sigma_{\text{vessels}} = 0.1 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$).

Impact of vessels on the treated volume

Table 2 shows the result of volume analysis for the 576 simulations (288 homogeneous and 288 non-homogeneous). The average treated volume (regions in which $|\vec{E}| > 700 \text{ V/cm}$) is indicated for the two conductivity models (homogeneous and heterogeneous), the six assayed locations and for every combination of electrodes separation distances, active lengths and voltages. In addition, out of the six assayed locations, the maximum relative difference between the two conductivity models is indicated.

Table 2. Average of treatment volumes (for six locations) and the maximum percentage deviation between homogeneous and non-homogeneous models.

Active Length (mm)	Space (mm)	3000 V			2500 V			2000 V		
		NHT	HT	NHT	NHT	HT	NHT	NHT	HT	NHT
		(cm ³)	(cm ³)	(%D)	(cm ³)	(cm ³)	(%D)	(cm ³)	(cm ³)	(%D)
10	10	3.98	4.00	1.27	3.19	3.22	1.59	2.39	2.42	2.55
	15	4.76	4.83	3.43	3.86	3.91	2.81	2.56	2.59	2.39
	20	5.06	5.05	0.48	3.49	3.49	2.15	2.10	2.09	2.74
	25	4.35	4.33	2.70	3.00	2.96	2.75	2.06	2.01	4.37
15	10	5.43	5.46	1.22	4.34	4.38	2.29	3.28	3.32	1.83
	15	6.50	6.47	5.20	6.22	6.29	2.78	3.81	3.83	2.55
	20	7.67	7.70	1.72	5.23	5.22	1.19	3.01	3.02	1.47
	25	6.56	6.54	1.47	4.32	4.29	0.92	2.62	2.61	0.87
20	10	7.04	7.05	0.45	5.66	5.68	0.84	4.28	4.28	0.86
	15	8.68	8.73	0.93	6.80	6.80	0.87	4.82	4.82	0.58
	20	9.31	9.51	3.69	6.68	6.85	4.05	3.88	3.96	2.62
	25	8.83	8.94	4.24	5.50	5.57	4.02	3.25	3.25	2.10
25	10	8.57	8.63	2.19	6.91	6.95	1.86	5.26	5.28	1.26
	15	10.07	10.05	4.05	8.19	8.22	2.17	6.12	6.13	2.79
	20	11.26	11.54	5.14	8.28	8.49	5.41	5.06	5.18	3.21
	25	10.99	11.17	2.62	6.94	7.10	6.04	3.88	3.90	4.84

Note: NHT: Non Homogeneous Tumor tissue, HT: Homogeneous Tumor tissue, (%D): maximum percentage deviation between the homogeneous and the non homogeneous cases.

The simulation results indicate that there are differences between homogeneous and non-homogeneous cases regarding the treated volume. It can be observed that the average integrated volume in cm^3 increases as electrode depth, spacing, and applied potential increase. The maximum percentage deviation of the tumor models ranged from 0.5% up to 6%.

Undertreatment volume quantification

The volume of tissue subjected to undertreatment because of the presence of blood vessels (heterogeneous case) was quantified by estimating the integrated volume inside the target region (regions in which $|\vec{E}| > 700 \text{ V/cm}$ for the homogeneous case) which had an electric field magnitude below 700V/cm. Results in Table 3 show the minimum, maximum and average undertreatment volumes in mm^3 for the various electrode configurations (15-25 mm depth, 10-25 mm spacing, and 3000-2000 V applied potential).

The results in Table 3 indicate that the minimum and maximum undertreatment volumes found for the 10, 15, 20 and 25 mm active lengths were (0- 68.7), (0-120.8), (0- 224.7) and (1.4- 227.7) mm^3 respectively. In some cases, in which the electrodes were inserted near very small vessels or $> 10 \text{ mm}$ away from vessels with diameters ranging from 2 to 8 mm, the undertreatment volume was close to 0 mm^3 .

Table 3. Volumes of tissue subjected to undertreatment because of the presence of blood vessels

Active length (mm)	Space (mm)	3000 V			2500 V			2000 V		
		Min (mm ³)	Max (mm ³)	Mean (mm ³)	Min (mm ³)	Max (mm ³)	Mean (mm ³)	Min (mm ³)	Max (mm ³)	Mean (mm ³)
10	10	0.0	31.8	19.5	19.2	35.2	28.3	0.0	26.6	14.4
	15	0.0	68.7	30.7	0.0	62.8	27.7	0.0	33.0	13.3
	20	1.6	6.3	4.1	0.0	7.0	4.1	0.0	11.9	4.1
	25	0.0	7.0	3.7	0.0	4.8	2.4	0.0	5.0	2.4
15	10	0.0	15.4	7.4	0.0	11.2	5.3	0.0	5.8	2.6
	15	61.9	120.8	97.3	21.3	81.9	51.4	0.0	30.4	20.1
	20	9.3	27.1	19.6	0.0	30.6	15.1	0.0	25.0	11.8
	25	1.7	22.8	10.8	0.0	7.9	2.9	0.0	4.5	1.7
20	10	0.0	43.3	18.4	0.0	28.0	14.5	0.0	20.9	12.9
	15	0.0	44.2	26.4	0.0	40.3	17.2	0.0	25.2	12.2
	20	62.7	224.7	111.4	50.4	201.7	96.8	24.4	112.1	61.5
	25	1.2	159.4	82.9	0.0	109.6	50.0	0.0	63.2	28.6
25	10	11.3	36.4	21.0	5.4	40.8	21.4	5.5	33.8	18.6
	15	5.6	119.5	47.5	1.8	65.9	38.3	1.4	38.4	17.9
	20	84.2	227.7	139.9	72.9	210.6	128.2	36.2	134.0	74.8
	25	74.5	182.6	112.2	27.4	155.3	66.6	11.3	50.2	27.3

Note: Min: Minimum, Max: Maximum.

Figure 3 illustrates two cases of undertreatment around the vessels. The green regions represent the isosurfaces of treatment ($|\vec{E}| = 700$ V/cm) for the homogeneous

model. The undertreatment regions ($|\vec{E}| < 700$ V/cm) for the heterogeneous models are depicted in blue.




Figure 3. 3D volume plot of vessels (red), homogeneous treatment region (green), tumor region (gray, only in b) and untreated spots (blue). (a) Deep treatment of liver (no tumor) close to the right hepatic vein and the portal vein with 25mm electrode active length and spacing, and 3000V applied potential. (b) Tumor case study close to vessels with 15mm active length, 14mm spacing and 2000V applied potential (case 2D illustrated in figure 2c).

The result depicted in Figure 3a (case H2) corresponds to the largest undertreatment volume found heuristically when very deep ablations were assayed. The numerical results corresponding to this case shown in figure 3a indicate that the undertreatment volume was 1049, 731.6, and 313.2 mm³ for the 3000, 2500, and 2000 V applied potentials respectively.

Simulated probability of undertreatment

The event rate of undertreatment was estimated from the 288 simulated cases, and it was considered to occur when there was at least an undertreated region ($|\vec{E}| < 700$ V/cm) with a volume ≥ 1 mm³ found within the IRE treatment region ($|\vec{E}| > 700$ V/cm for the homogeneous case).

Table 4. Number of cases of undertreatment (undertreated volume ≥ 1 mm³) out of the 6 assayed insertion locations

Electric Potential(V)	Active length (mm)	Space(mm)			
		10	15	20	25
2000	10	2/6	2/6	2/6	3/6
	15	2/6	2/6	3/6	3/6
	20	3/6	4/6	4/6	4/6
	25	3/6	4/6	4/6	5/6
2500	10	2/6	2/6	3/6	3/6
	15	2/6	3/6	3/6	4/6
	20	4/6	4/6	5/6	6/6

	25	5/6	5/6	6/6	6/6
	10	3/6	2/6	4/6	4/6
	15	3/6	4/6	5/6	6/6
3000	20	4/6	5/6	5/6	6/6
	25	5/6	5/6	6/6	6/6

The results in Table 4 indicate that the minimum and maximum probability of having undertreatment volume ($\geq 1 \text{ mm}^3$) found for the 10mm active length electrode was (2/6 - 4/6) and for the 15, 20 and 25 mm active lengths electrodes were (2/6 - 6/6).

Furthermore, it was found that the presence of vessels with a diameter $\geq 3\text{mm}$ within the target region always produced at least a region of undertreatment with a volume $\geq 1 \text{ mm}^3$.

Three and Four-Electrode Arrays

In order to be further relevant to clinical treatment, we simulated three-electrode and four-electrode arrays with the various electrode configurations (20-25 mm depth and 15-25 mm spacing). This was performed by applying 2000, 2500, and 3000V to the possible electrode sets and between the two diagonal pairs.

The result depicted in Figure 4 shows four possible treatment planning approaches for the targeted tumor in case H1 depicted in Figure 2. Figure 4a shows the result of a 2-

electrode model with 15mm active length. The distortion on the electric field distribution is clearly observed around the blood vessels. Figure 4b shows the result of four-electrode array simulation with 15mm active length. It can be observed that the four-electrode array increased the treatment zone and reduced the undertreatment region, however, a small area of electric field distortion was still noticed around the vessel. Figure 4c shows the simulation result of three-electrode array simulation with 20mm active length. It can be observed that the increase of active length increased the treatment region, nonetheless, a small region of electric field distortion is also clearly noticed around the vessels. The result depicted in Figure 4d shows the result of low conductivity liquid infusion (0.1S/m); see next section. A uniform electric field distribution was produced and it was noticed that the undertreatment region around the vessels was completely covered.




Figure 4. X-Y plane plot of electric field magnitude during hypothetical treatment of a tumor. (a) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) two 15mm electrodes model result. (b) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) four 15mm electrodes model result. (c) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) three 20 mm electrodes model result. (d) Result of two 15 mm electrodes model when the vessel is filled with a low conductivity fluid ($\sigma_{\text{vessels}} = 0.1 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$).

Furthermore, we repeated the simulation of the maximum undertreatment volumes found in our study with three-electrode and four-electrode arrays in order to assess the impact of these approaches on the undertreatment volumes. The results depicted in Table 5 shows the maximum undertreatment volume found in the repeated simulations. The maximum undertreatment volumes found for the 20 and 25 mm active lengths were (227.7), (291.9) and (215.7) mm^3 for 2, 3 and 4 electrode arrays respectively. In cases where the electrodes were inserted near a heavy set of vessels, a larger undertreatment volume was sometimes found.

Table 5. Volumes of tissue subjected to undertreatment because of the presence of blood vessels of different treatment approaches

Active length (mm)	Spacing (mm)	Potential (V)	Single Active Probe Max (mm^3)	Three Active Probe Max (mm^3)	Two Probe Max (mm^3)	Saline Infusion (0.1S/m) Max (mm^3)
20	20	2000	112.1	108.1	43.9	0
		2500	201.7	160.4	87.5	0
		3000	224.7	254.3	160.3	2.1
	25	2000	63.2	54.2	15.1	0
		2500	109.6	87.8	32.2	0
		3000	159.4	142.6	66.7	0

25	20	2000	134.0	68.1	32.7	0
		2500	210.6	157.4	48.2	0
		3000	227.7	291.9	215.7	0
	25	2000	50.2	80.1	32	1.1
		2500	155.3	118.5	28.9	1.8
		3000	182.6	148.2	39.6	2.4

Simulated potential benefit of infusing low conductivity fluids into liver vasculature for preventing undertreatment

In order to prevent undertreatment due to the heterogeneity conductivity caused by the liver vessels, we propose to inject the blood vessels with a low conductivity liquid so that the tissues become more homogeneous in terms of conductivity. Figure 2d illustrates that the electric field distribution obtained with a model in which the vessel has a conductivity of 0.1 S/m is almost indistinguishable when compared to that obtained with the homogenous model. More importantly, the region of tumor undertreatment observed in Figure 2c disappears in Figure 2d.

The results in Table 6 indicate that by reducing the modeled conductivity of blood vessels, the size of the undertreated volumes is extremely reduced and in most cases, undertreatment is not noticeable at all. The event rate of undertreatment is reported in Table 7 when the conductivity of the vessels is set to 0.1 S/m rather than 0.7 S/m.

Table 6. Average undertreated volume (mm³) for different electrical conductivities of the liver vessels

Active length (mm)	Space (mm)	3000V		2500V		2000V	
		0.7	0.1	0.7	0.1	0.7	0.1
		(S/m)	(S/m)	(S/m)	(S/m)	(S/m)	(S/m)
10	10	31.8	0.0	35.2	0	26.6	0
	15	68.7	0	62.8	0	33.0	0
	20	6.3	0	7.0	0	11.9	0
	25	7.0	3.7	4.8	1.6	5.0	2.5
15	10	15.4	0	11.2	0	5.8	0
	15	120.8	0	81.9	0	30.4	0
	20	27.1	2.8	30.6	0	25.0	0
	25	22.8	0	7.9	0	4.5	0
20	10	43.3	0	28.0	0	20.9	0
	15	44.2	0	40.3	0	25.2	0
	20	224.7	2.1	201.7	0	112.1	0
	25	159.4	0	109.6	0	63.2	0
25	10	36.4	0	40.8	0	33.8	0
	15	119.5	0	65.9	0	38.2	0
	20	227.7	0	210.6	0	134.0	0
	25	182.6	2.4	155.3	1.8	50.2	1.1

Table 7. Number of cases of undertreatment (undertreated volume

$\geq 1\text{mm}^3$) out of the 6 assayed insertion locations when the conductivity of the vessels is 0.1 S/m.

Electric Potential(V)	Active length (mm)	Space(mm)			
		10	15	20	25
2000	10	0/6	0/6	0/6	2/6
	15	0/6	0/6	0/6	0/6
	20	0/6	0/6	0/6	0/6
	25	0/6	0/6	0/6	1/6
2500	10	0/6	0/6	0/6	1/6
	15	0/6	0/6	0/6	0/6
	20	0/6	0/6	0/6	0/6
	25	0/6	0/6	0/6	0/6
3000	10	0/6	0/6	0/6	1/6
	15	0/6	0/6	2/6	0/6
	20	0/6	0/6	2/6	0/6
	25	0/6	2/6	2/6	2/6

Discussion

Complete ablation of viable tumor cells by achieving an

adequate electric field distribution is the purpose of IRE treatment planning. Thus, understanding the impact of the treatment area characteristics, the electric conductivity heterogeneities in particular, on the electric field distribution is crucial to achieving optimal treatment. In this study, the impact of blood vessels in the treatment zone was examined in a realistic anatomical model consisting of liver, blood vessels, bones and surrounding tissues which were obtained from human male CT imaging. In earlier studies, it has been found that large blood vessels, because of their high conductivity, may interfere with electric field coverage of tumors in liver (14-16). Golberg et al. (14) described the impact of blood vessels on electric field distribution of IRE treatment. In

which they experimentally demonstrated the need to increase the baseline field intensity around large vessels. Unsurprisingly, our study found that the distortion produced by the conductivity heterogeneity caused by the presence of a blood vessel was dependant on the vessel diameter and on the electrode's distance to the vessel. This is in agreement with other previous studies (14-16, 27-29).

In the present study, we have modeled the blood vessels as simple containers of blood. That is, they were modeled as homogeneous regions of higher electrical conductivity than the liver parenchyma. Unfortunately, our data model didn't support the inclusion of vessels wall. Furthermore, the thickness of vessel walls varies enormously around arteries and veins let alone variability in the structure between veins and arteries. The lack of existing data about vessel walls as well as the non uniform geometry distribution of vessels limited the possibility of adding accurate vessel walls around our data model. However, to validate this simplification, an additional group of simulations was repeated with the inclusion of a 1 mm vessel wall around large hepatic veins and portal vein, as described in our previous study (30). The results of these simulations is included into the supplementary materials (Figure1 and table 7). The results show that the presence of the wall reduces the influence of blood on the electric field distribution and the undertreatment volumes. However, the relative difference in undertreatment volume compared to the results reported here does not exceed 21% .

In terms of treatment volume size (Table 2), our results indicate that the impact of blood vessels will be mild for superficial liver IRE treatments. The maximum deviation in volume we observed was only 6%, which can be perceived as almost negligible if other confounding factors that influence the outcome of IRE are considered. For instance,

tolerances when placing the electrodes are likely to cause a larger error in treatment planning. On the other hand, the observed regions of undertreatment around the blood vessels represent a potentially grave issue. We observed undertreatment in 65% of the analyzed cases (Table 4). This high incidence rate would increase if deeper treatments – such as the ones represented in Figure 3 – were analyzed because the electrodes would be closer to even larger vessels. Therefore, we deem that undertreatment is likely to occur in current IRE liver treatments. It must be noted that IRE is particularly prescribed for ablations close to vessels because in those scenarios other ablative techniques based on temperature are not usable (31). Whether undertreated regions of only 1 mm³ represent a real clinical threat is still debatable, particularly because in our analysis undertreatment typically occurred at the periphery of the target volume and it must be presumed that in clinical IRE these locations would correspond to the safety margins considered during treatment planning. However, undertreatment certainly should be a matter of concern because any tumor cell left viable is a new potential tumor nodule. Furthermore, it must also be taken into account that we obtained much larger volumes of undertreatment, reaching up to 228 mm³ (Table 3). Therefore, although to the best of our knowledge no clinical cases of recurrence after hepatic IRE have been reported that were directly linked to undertreatment around vessels, we advise practitioners of liver IRE to carefully monitor any possible signs of recurrence around vessels during post-treatment follow-ups. Again it is convenient to note that undertreatment around vessels has been experimentally reported by Golberg et al. (14).

Different treatment planning approaches were included in our simulation in order to be more relevant to clinical applications. Three-electrode and four-electrode arrays were

simulated. Our results show an increase in the treatment zones and a decrease in the size of undertreatment volumes (Table 5). The four-electrode array produced a larger treatment zone similar to the reported results in Applebaum et al. (11).

In order to prevent undertreatment due to the presence of vessels, we propose to replace blood in the liver vessels with an isotonic low conductivity liquid – such as a mixture of normal saline and dextrose 5% – so that the tissues are made more homogeneous in terms of conductivity. Our results indeed showed a huge improvement regarding cases of undertreatment (Table 7) and regarding the size of the undertreated regions (Table 6) when the conductivity of the vessels is set to 0.1 S/m. Similar results, not reported here, were obtained with higher (0.2 S/m) and lower (0.05 S/m) conductivities.

Delivery of low conductivity liquids into the liver vasculature could be performed either directly through percutaneous transhepatic portal catheterization (PTPC), which has been used for islet transplantation (32), or through hepatic vascular exclusion (HVE) maneuvers by deploying clamps through laparoscopy (33). It is worth noting that total HVE (THVE) has been used for gene therapy (34) or to improve the pharmacokinetics of chemotherapeutic drugs (35). Several low conductivity solutions could be employed with a reasonable margin of safety (36-38). In particular, by diluting one part of normal saline (0.9% NaCl, $\sigma_{37^{\circ}\text{C}} = 1.5 \text{ S/m}$) in fourteen parts of dextrose 5% ($\sigma_{37^{\circ}\text{C}} < 0.01 \text{ S/m}$) an isotonic solution with a conductivity of about 0.1 S/m would be obtained. Note that, if the delivery technique does not substitute the blood but merely dilutes it, then the conductivity of the injected liquid would have to be adjusted to a lower value so that the resulting mixture had an adequate conductivity (0.1 S/m).

Conclusion

Due to their impact on electric field distribution, liver blood vessels may have an impact on IRE treatments in terms of treatment volume size and shape. This numerical study shows that the impact of blood vessels occurs around the medium size vessels and on the periphery of the treatment region, which may sufficiently distort the electric field distribution that resulted in undertreatment regions, with the potential effect on the quality of IRE cancer treatment. According to the reported results, the frequent random positions with different electrode settings indicated that the probability of undertreatment could occur frequently in hepatic IRE treatments. These undertreatment regions correspond to areas where the applied electric field is potentially sub-lethal. Therefore, caution is advised due to the possibility of these undertreated volumes to become new tumor nodules. This risk could be minimized by replacing blood in the liver vasculature, or by diluting it, with isotonic low conductivity liquids.

Declaration of Conflicting Interests

The authors report no potential conflicts of interest.

Acknowledgements

The authors would like to thank the Korea Institute of Science and Technology Information (KISTI) for their assistance in providing us with the anatomical geometry.

Funding

This work received financial support from the Spanish Government through grants TEC2010-17285, TEC2011-27133-C02 and TEC2014-52383-C3 (TEC2014-52383-C3-2) and from the European Commission through the Marie Curie IRG grant “TAMIVIVE” (256376).

References

1. Rubinsky B, Onik G, Mikus P. Irreversible electroporation: a new ablation modality--clinical implications. *Technol Cancer Res Treat*. 2007;6(1):37-48. doi:d=3029&c=4222&p=15542&do=detail [pii].
2. Davalos R V., Mir LM, Rubinsky B. Tissue ablation with irreversible electroporation. *Ann Biomed Eng*. 2005;33(2):223-231. doi:10.1007/s10439-005-8981-8.
3. Onik G, Rubinsky B. Irreversible Electroporation: First Patient Experience Focal Therapy of Prostate Cancer. In: Rubinsky B, ed. *Irreversible Electroporation SE - 10*. Series in Biomedical Engineering. Springer Berlin Heidelberg; 2010:235-247. doi:10.1007/978-3-642-05420-4_10.
4. Thomson KR, Cheung W, Ellis SJ, et al. Investigation of the Safety of Irreversible Electroporation in Humans. *J Vasc Interv Radiol*. 2011;22(5):611-621. doi:http://dx.doi.org/10.1016/j.jvir.2010.12.014.
5. Pech M, Janitzky A, Wendler J, et al. Irreversible Electroporation of Renal Cell Carcinoma: A First-in-Man Phase I Clinical Study. *Cardiovasc Intervent Radiol*. 2011;34(1):132-138. doi:10.1007/s00270-010-9964-1.
6. Philips P, Hays D, Martin RCG. Irreversible electroporation ablation (IRE) of unresectable soft tissue tumors: Learning curve evaluation in the first 150 patients treated. *PLoS One*. 2013;8(11):1-9. doi:10.1371/journal.pone.0076260.
7. Narayanan G, Bhatia S, Echenique A, Suthar R, Barbery K, Yrizarry J. Vessel Patency Post Irreversible Electroporation. *Cardiovasc Intervent Radiol*. 2014;37(6):1523-1529. doi:10.1007/s00270-014-0988-9.
8. Kingham TP, Karkar AM, D'Angelica MI, et al. Ablation of perivascular hepatic malignant tumors with irreversible electroporation. *J Am Coll Surg*. 2012;215(3):379-387. doi:10.1016/j.jamcollsurg.2012.04.029.
9. Lee YJ, Lu DSK, Osuagwu F, Lassman C. Irreversible Electroporation in Porcine Liver: Short- and Long-Term Effect on the Hepatic Veins and Adjacent Tissue by CT With Pathological Correlation. *Invest Radiol*. 2012;47(11). Available at: http://journals.lww.com/investigativeradiology/Fulltext/2012/11000/Irreversible_Electroporation_in_Porcine_Liver_.8.aspx.
10. Maor E, Ivorra A, Leor J, Rubinsky B. The effect of irreversible electroporation on blood vessels. *Technol Cancer Res Treat*. 2007;6(4):307-312.
11. Appelbaum L, Ben-David E, Faroja M, Nissenbaum Y, Sosna J, Goldberg S. Irreversible electroporation ablation : Creation of large-volume ablation zones in in vivo porcine liver with four-electrode arrays. *Radiology*. 2014;270(2):416-424. doi:10.1148/radiol.13130349.
12. Lee EW, Chen C, Prieto VE, Dry SM, Loh CT, Kee ST. Advanced hepatic ablation technique for creating complete cell death: irreversible electroporation. *Radiology*. 2010;255(2):426-433. doi:10.1148/radiol.10090337.

13. Ivorra A. Tissue Electroporation as a Bioelectric Phenomenon: Basic Concepts. In: Rubinsky B, ed. *Irreversible Electroporation*. Springer Berlin Heidelberg; 2010:23-61. doi:10.1007/978-3-642-05420-4_2.
14. Golberg A, Bruinsma BG, Uygun BE, Yarmush ML. Tissue heterogeneity in structure and conductivity contribute to cell survival during irreversible electroporation ablation by “electric field sinks.” *Sci Rep*. 2015;5:8485. doi:10.1038/srep08485.
15. Marčan M, Kos B, Miklavčič D. Effect of Blood Vessel Segmentation on the Outcome of Electroporation-Based Treatments of Liver Tumors. *PLoS One*. 2015;10(5):e0125591. doi:10.1371/journal.pone.0125591.
16. Qasrawi R, Ivorra A. Impact of Liver Vasculature on Electric Field Distribution during Electroporation Treatments: An Anatomically Realistic Numerical Study. In: Lacković I, Vasic D, eds. *6th European Conference of the International Federation for Medical and Biological Engineering SE - 143*. Vol 45. IFMBE Proceedings. Springer International Publishing; 2015:573-576. doi:10.1007/978-3-319-11128-5_143.
17. Ben-David E, Ahmed M, Faroja M, et al. Irreversible Electroporation: Treatment Effect Is Susceptible to Local Environment and Tissue Properties. *Radiology*. 2013;269(3):738-747. doi:10.1148/radiol.13122590.
18. Shin DS, Chung MS, Lee JW, et al. Advanced surface reconstruction technique to build detailed surface models of the liver and neighboring structures from the Visible Korean Human. *J Korean Med Sci*. 2009;24(3):375-383. doi:10.3346/jkms.2009.24.3.375.
19. Larrabide I, Omedas P, Martelli Y, et al. GIMIAS: An open source framework for efficient development of research tools and clinical prototypes. *Lect Notes Comput Sci (including Subser Lect Notes Artif Intell Lect Notes Bioinformatics)*. 2009;5528:417-426. doi:10.1007/978-3-642-01932-6_45.
20. Mekid H, Mir LM. A validated model of in vivo electric field distribution in tissues for electrochemotherapy and for DNA electrotransfer for gene therapy. 2000;1523:73-83.
21. Guo Y, Zhang Y, Klein R, et al. Irreversible electroporation therapy in the liver: Longitudinal efficacy studies in a rat model of hepatocellular carcinoma. *Cancer Res*. 2010;70(4):1555-1563. doi:10.1158/0008-5472.CAN-09-3067.
22. Ivorra A, Mir LM, Rubinsky B. Electric Field Redistribution due to Conductivity Changes during Tissue Electroporation: Experiments with a Simple Vegetal Model. *IFMBE Proc*. 2009;25(13):59-62.
23. Neal RE, Garcia P a., Robertson JL, Davalos R V. Experimental characterization and numerical modeling of tissue electrical conductivity during pulsed electric fields for irreversible electroporation treatment planning. *IEEE Trans Biomed Eng*. 2012;59(4):1076-1085. doi:10.1109/TBME.2012.2182994.

24. Ivorra A, Rubinsky B. In vivo electrical impedance measurements during and after electroporation of rat liver. *Bioelectrochemistry*. 2007;70(2):287-295. doi:10.1016/j.bioelechem.2006.10.005.
25. Prakash S, Karnes MP, Sequin EK, et al. Ex vivo electrical impedance measurements on excised hepatic tissue from human patients with metastatic colorectal cancer. *Physiol Meas*. 2015;36(2):315-328. doi:10.1088/0967-3334/36/2/315.
26. Gabriel S, Lau RW, Gabriel C. The dielectric properties of biological tissues: II. Measurements in the frequency range 10 Hz to 20 GHz. *Phys Med Biol*. 1996;41(11):2251-2269. doi:10.1088/0031-9155/41/11/002.
27. Charpentier KP, Wolf F, Noble L, Winn B, Resnick M, Dupuy DE. Irreversible electroporation of the liver and liver hilum in swine. *Hpb*. 2011;13(3):168-173. doi:10.1111/j.1477-2574.2010.00261.x.
28. Marcan M, Pavliha D, Music MM, Fuckan I, Magjarevic R, Miklavcic D. Segmentation of Hepatic Vessels from Mri Images for Planning of Electroporation-Based Treatments in the Liver. *Radiol Oncol*. 2014:1-15. doi:10.2478/raon-2014-0022.
29. Marčan M, Kos B, Miklavčič D. Effect of Blood Vessel Segmentation on the Outcome of Electrochemotherapy of Liver Tumors - Preliminary Theoretical Study. In: Lacković I, Vasic D, eds. *6th European Conference of the International Federation for Medical and Biological Engineering SE - 205*. Vol 45. IFMBE Proceedings. Springer International Publishing; 2015:825-828. doi:10.1007/978-3-319-11128-5_205.
30. Silve L, Qasrawi R, Ivorra A. Incorporation of the Blood Vessel Wall into Electroporation Simulations. In: 1st World Congress on Electroporation and Pulsed Electric Fields in Biology, Medicine and Food & Environmental Technologies. Springer; 2016:223-227.
31. Scheffer HJ, Nielsen K, de Jong MC, et al. Irreversible electroporation for nonthermal tumor ablation in the clinical setting: a systematic review of safety and efficacy. *J Vasc Interv Radiol*. 2014;25(7):997-1011.
32. Gao F, Ai S-D, Liu S, Zeng W-B, Wang W. Percutaneous transhepatic portal catheterization guided by ultrasound technology for islet transplantation in rhesus monkey. *Hepatobiliary Pancreat Dis Int*. 2012;11(2):154-159.
33. Lau W-Y, Lai EC, Lau SH. Methods of vascular control technique during liver resection: a comprehensive review. *Hepatobiliary Pancreat Dis Int*. 2010;9(5):473-481.
34. Alves A, Adam R, Majno P, et al. Hepatic resection for metastatic renal tumors: is it worthwhile? *Ann Surg Oncol*. 2003;10(6):705-710.

35. Averbach AM, Stuart OA, Sugarbaker TA, et al. Pharmacokinetic studies of intraaortic stop-flow infusion with 14 C-labeled mitomycin C. *J Surg Res.* 1995;59(3):415-419.
36. Wolf A V. Aqueous solutions and body fluids : their concentrative properties and conversion tables. New York; London: Hoeber Medical Division Harper & Row Publ; 1966.
37. Gananadha S, Morris DL. Saline Infusion Markedly Reduces Impedance and Improves Efficacy of Pulmonary Radiofrequency Ablation. *Cardiovasc Intervent Radiol.* 2004;27(4). doi:10.1007/s00270-003-0112-z.
38. Gillams AR, Lees WR. CT mapping of the distribution of saline during radiofrequency ablation with perfusion electrodes. *Cardiovasc Intervent Radiol.* 2005;28(4):476-480.

SUPPLEMENTARY MATERIALS

Table 1. Average of treatment volumes and the maximum percentage deviation between homogeneous and non-homogeneous models. Considering in the model the electrical conductivity of healthy (normal) liver.

Active Length (mm)	Space (mm)	3000 V			2500 V			2000 V		
		NHN	HN	NHN	NHN	HN	NHN	NHN	HN	NHN
		(cm ³)	(cm ³)	(%D)	(cm ³)	(cm ³)	(%D)	(cm ³)	(cm ³)	(%D)
10	10	4.7	4.8	2.3	3.7	3.8	3.0	2.8	2.9	3.3
	15	5.8	5.9	5.1	4.4	4.5	4.8	3.1	3.1	2.3
	20	6.5	6.5	1.0	4.8	4.8	0.8	2.9	2.9	1.1
	25	6.4	6.5	6.8	4.0	4.1	7.3	2.3	2.3	3.2
15	10	6.3	6.3	2.5	5.0	5.1	3.0	3.8	3.8	3.0
	15	14.2	14.4	2.4	9.6	9.7	2.9	4.5	4.4	3.9
	20	9.2	9.3	1.0	6.9	6.9	1.2	4.3	4.3	2.1
	25	9.0	9.1	0.8	6.0	5.9	1.3	3.2	3.1	3.6
20	10	8.0	8.0	0.9	6.4	6.4	0.7	4.9	4.9	1.4
	15	10.1	10.1	1.0	8.0	8.0	1.6	5.7	5.7	1.5
	20	11.1	11.4	5.6	8.2	8.5	4.7	5.1	5.2	5.6
	25	11.8	11.9	6.5	7.8	7.9	6.5	3.8	3.7	6.8
25	10	9.7	9.7	2.6	7.8	7.8	1.9	5.9	5.9	1.3
	15	12.0	12.0	1.8	9.5	9.5	1.6	6.8	6.8	0.9
	20	13.6	13.9	2.8	10.2	10.4	3.1	6.5	6.6	3.0
	25	14.3	14.5	2.1	9.6	9.7	2.6	4.5	4.4	3.3

Note: NHN: Non Homogeneous Normal tissue, HN: Homogeneous Normal tissue, (%D): maximum percentage deviation between the homogeneous and the non homogeneous cases.

Table 2. Volumes of tissue subjected to undertreatment because of the presence of blood vessels. Considering in the model the electrical conductivity of healthy (normal) liver.

Active length (mm)	Space (mm)	3000 V			2500 V			2000 V		
		Min (mm ³)	Max (mm ³)	Mean (mm ³)	Min (mm ³)	Max (mm ³)	Mean (mm ³)	Min (mm ³)	Max (mm ³)	Mean (mm ³)
10	10	0.0	53.3	34.5	0.0	55.1	35.1	0.0	48.7	30.1
	15	0.0	119.1	58.6	0.0	82.4	40.6	0.0	33.1	19.2
	20	4.4	10.1	7.9	1.6	5.9	4.4	0.0	4.9	3.2
	25	3.0	12.2	7.8	0.0	8.2	3.7	0.0	6.5	3.3
15	10	0.0	28.6	13.8	0.0	17.3	8.5	0.0	13.2	6.3
	15	114.6	160.6	145.2	93.8	109.1	103.2	14.6	41.8	30.4
	20	15.6	54.7	37.0	6.1	53.1	31.7	0.0	48.3	21.5
	25	4.2	45.4	20.4	2.0	24.8	10.8	0.0	8.2	3.1
20	10	2.0	76.2	32.7	1.3	52.9	27.0	0.0	37.5	20.4
	15	0.0	109.4	51.8	0.0	62.1	34.6	0.0	53.4	28.4
	20	120.1	443.7	216.7	90.2	370.0	171.8	41.2	193.2	113.4
	25	49.9	291.6	172.1	35.1	224.4	142.3	0.0	88.0	39.8
25	10	31.1	63.9	48.1	16.6	54.2	37.6	9.2	41.8	27.8
	15	9.5	121.5	55.4	4.1	76.5	34.7	1.5	41.7	18.1
	20	114.8	402.0	212.6	97.9	344.6	176.9	57.2	200.6	103.2
	25	74.5	197.9	143.6	60.8	161.9	99.9	19.5	52.0	32.4

Note: Min: Minimum, Max: Maximum.

Table 3. **Number of cases of undertreatment (undertreated volume $\geq 1\text{mm}^3$) out of the 6 assayed insertion locations. Considering in the model the electrical conductivity of healthy (normal) liver.**

Electric Potential(V)	Active length (mm)	Space(mm)			
		10	15	20	25
2000	10	2/6	2/6	2/6	3/6
	15	2/6	2/6	3/6	3/6
	20	3/6	4/6	4/6	4/6
	25	3/6	4/6	4/6	5/6
2500	10	2/6	2/6	3/6	3/6
	15	2/6	3/6	4/6	4/6
	20	4/6	5/6	6/6	6/6
	25	5/6	5/6	6/6	6/6
3000	10	3/6	4/6	4/6	4/6
	15	3/6	4/6	5/6	6/6
	20	4/6	5/6	5/6	6/6
	25	5/6	5/6	6/6	6/6

Table 4. **Average undertreated volume (mm^3) for different electrical conductivities of the liver vessels. . Considering in the model the electrical conductivity of healthy (normal) liver.**

Active length (mm)	Space (mm)	3000V		2500V		2000V	
		0.7	0.1	0.7	0.1	0.7	0.1
		(S/m)	(S/m)	(S/m)	(S/m)	(S/m)	(S/m)
10	10	34.5	0.0	35.1	0	30.1	0
	15	58.6	0	40.6	0	19.2	0
	20	7.9	0	4.4	0	3.2	0

15	25	7.8	6.5	3.7	3.9	3.3	3.2
	10	13.8	0	8.5	0	6.3	0
	15	145.2	1.1	103.2	1.4	30.4	1.2
	20	37.0	1.2	31.7	0	21.5	3.6
	25	20.4	2.7	10.8	0	3.1	0
20	10	32.7	0	27.0	0	20.4	0
	15	51.8	0	34.6	0	28.4	0
	20	216.7	3.3	171.8	1.5	113.4	1.2
	25	172.1	7.9	142.3	17.7	39.8	3.1
	10	48.1	2	37.6	1	27.8	0.7
25	15	55.4	7.3	34.7	4.9	18.1	1.4
	20	212.6	11	176.9	6	103.2	2
	25	143.6	3.2	99.9	2.8	32.4	1.5

Table 5. Number of cases of undertreatment (undertreated volume $\geq 1\text{mm}^3$) out of the 6 assayed insertion locations When the conductivity of the vessels is 0.1 S/m. Considering in the model the electrical conductivity of healthy (normal) liver.

Electric Potential(V)	Active length (mm)	Space(mm)			
		10	15	20	25
2000	10	0/6	0/6	0/6	2/6
	15	0/6	1/6	2/6	0/6
	20	0/6	0/6	2/6	2/6
	25	1/6	2/6	2/6	2/6
2500	10	0/6	0/6	0/6	2/6

	15	0/6	2/6	0/6	0/6
	20	0/6	0/6	1/6	3/6
	25	1/6	3/6	3/6	1/6
Table 7 Volumes of tissue subjected to undertreatment because of the presence of blood vessels with and without the inclusion of vessel wall					
	10	0/6	0/6	0/6	2/6
	15	0/6	2/6	2/6	3/6
3000	20	0/6	0/6	0/6	2/6
	25	0/6	1/6	2/6	0/6

Table 6 . **Volumes of tumor tissue subjected to undertreatment because of the presence of blood vessels of different IRE thresholds (V/cm).**


Active length (mm)	Space (mm)	500 V/cm	600 V/cm	700 V/cm	800 V/cm	1000 V/cm
	15	95.7	63.1	44.2	35.2	15.2
20	20	319	272.3	224.8	180	49.2
	25	317.9	202.9	159.4	75.5	19.6
	15	190.7	142.1	119.5	87.8	41.1
25	20	273	271.2	227.7	206.3	98.4
	25	241.5	204.2	182.6	153.7	119.8

Note: 3000 V applied potential


Active length (mm)	Space (mm)	3000V			2500V			2000V		
		Max (mm ³)	Max-W (mm ³)	(%D)	Max (mm ³)	Max-W (mm ³)	(%D)	Max (mm ³)	Max-W (mm ³)	(%D)
15	15	120.8	99.1	18	81.9	66.3	19	30.4	27.1	11
	20	27.1	23.8	12	30.6	25.1	18	25	21.3	15
	25	22.8	18.9	17	7.9	6.8	14	4.5	3.6	19
20	15	44.2	37.1	16	40.3	32.6	19	25.2	22.2	12
	20	224.7	193.2	14	201.7	168.4	16.5	112.1	90.8	19
	25	159.4	135.5	15	109.6	91.0	17	63.2	52.5	17
	15	119.5	96.8	19	65.9	54.7	17	38.4	30.3	21
25	20	227.7	191.3	16	210.6	172.7	18	134	112.6	16
	25	182.6	160.7	12	155.3	132.0	15	50.2	41.2	18

Note: Max: Maximum undertreatment without vessel wall, Max-W: Maximum undertreatment with the vessel wall.


(%D): maximum percentage deviation between the blood vessels cases and the blood vessels and vessels wall cases.




a) Undertreatment volume = 159.4 mm³



b) Undertreatment volume = 132.1 mm³



c) Undertreatment volume = 93.2 mm³



d) Undertreatment volume = 0 mm³

Figure 1. X-Y plane plot of electric field magnitude during hypothetical treatment of a tumor. (a) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} = \sigma_{\text{wall}} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) result. (d) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} \neq \sigma_{\text{wall}} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) result (c) The non-homogeneous model ($\sigma_{\text{vessels}} = 0.7 \text{ S/m} \neq \sigma_{\text{wall}} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) result 3-electrode array (d) Result when the vessel is filled with a low conductivity fluid ($\sigma_{\text{vessels}} = 0.1 \text{ S/m} \neq \sigma_{\text{liver}} \neq \sigma_{\text{tumor}}$) .